

Test Report

Testing laboratory:

Nordic Institute of Dental Materials
Sognsveien 70 A, P.O. Box 3874 Ullevål stadion, NO-0805 Oslo, Norway

Test Report no. 017/13 issued: 06.08.2013

Client: DNV Nemko Presafe AS
Gaustadalléen 30
P.O. Box 73 Blindern
0373 Oslo Norway

Name of product: M&R Claushuis mercury

Batch number(s): not given

Testing contract number: T-017/13
Date of receipt of test item: 23.07.2013
Date(s) of performance of test(s): 06.08.2013 – 06.08.2013

Test methods and procedures according to:

ISO 24234:2004 Dentistry - Mercury and alloys for dental amalgam

For the specific results, see the following pages. This test report applies only to the items tested. This report shall not be reproduced, except in full, without written approval by NIOM.



Hilde Molvig Kopperud
Head of laboratory

ISO 24234:2004 Dentistry - Mercury and alloys for dental amalgam

Product: Mercury

Manufacturer: M&R Claushuis

The numbers below refer to the corresponding sections in the reference standard.

4 Requirements

4.5 Contamination of the mercury by oil, water and foreign material

Test method: 6.3.1, 6.3.2

Test results: The mercury is free from contamination and has a bright mirror-like surface.

Comments: The material complies with the requirements.

4.6 Purity of mercury- Free pouring

Test method: 6.4

Test results: The mercury pours freely without tailing.

Comments: The material complies with the requirements.

Technical responsibility



Liv Christiansen



Signature to verify that the data in this test report have been copied correctly from the laboratory journal